



WASHINGTON STATE PATROL  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Mar 7, 2006 INITIALS: DEC OWNER: WSP  
DISTRICT: 4 ADDRESS: Highway 395 MP 239  
LOCATION: Kettle Falls 7067 COUNTY: Stevens  
MAKE: Rice Lake TYPE: truck scale ACCURACY CLASS: III  
MODEL: TD 310 ID. NO.: 00160499 CAPACITY: 100,000 lb SECTIONS: 2  
MINIMUM GRADUATIONS: 20 lb REG. ELEMENT: digital PLATFORM: 10' X 14'

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
<u>Dist</u>	<u>4000</u>	<u>4000</u>	<u>0</u>	<u>Dist</u>	<u>36000</u>	<u>35980</u>	<u>-20</u>
<u>"</u>	<u>8000</u>	<u>8000</u>	<u>0</u>	<u>"</u>	<u>40000</u>	<u>39980</u>	<u>-20</u>
<u>"</u>	<u>12000</u>	<u>12000</u>	<u>0</u>				
<u>"</u>	<u>16000</u>	<u>16000</u>	<u>0</u>				
<u>"</u>	<u>20000</u>	<u>19980</u>	<u>-20</u>				
<u>"</u>	<u>24000</u>	<u>23980</u>	<u>-20</u>				
<u>"</u>	<u>28000</u>	<u>27980</u>	<u>-20</u>				
<u>Dist</u>	<u>32000</u>	<u>31980</u>	<u>-20</u>				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. L 2441-142

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Crona

Date: 3/9/06

State of Washington  
County of Thurston  
Signed and sworn to (or affirmed) before  
me on March 9, 2006 by

(Signature of notary public)

Printed Name of Notary

My appointment expires 12-27-07



3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS